

# M K I S



Calibration Company

Milton Keynes Instrumentation Services  
Unit 4 Furtho Court  
Towcester Road  
Old Stratford  
Milton Keynes  
MK19 6AQ

Tel: 01908 568250  
Fax: 01908 564661

## CERTIFICATE OF CALIBRATION

CERTIFICATE NO: 31811

DATE OF ISSUE: 16.04.07

<b>Customer:</b>	DJB Labcare Limited	<b>Order Number:</b>	Verbal
<b>Location:</b>	Newport Pagnell	<b>Serial Number:</b>	05000509
<b>Instrument Type:</b>	Omega HH506 RA	<b>Adjustment Required:</b>	No

All measurements were made using equipment whose values are referenced directly or by approved radiometric procedures to laboratory standards whose values are traceable to National Standards.

The ambient temperature and relative humidity throughout the test was  $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$  and  $50\% \pm 10$  respectively.

The calibration was carried out in accordance with the general requirements of BS EN ISO/IEC 17025: 2005.

This is to certify that the above listed instrument meets or exceeds as defined by our established calibration procedures, the published manufacturers specification.

The uncertainties refer to the applied and measured values only with no account taken of the instrument's ability to maintain its calibration.

0  $\mu\text{V}$  to 100 mV  $\pm$  (20 ppm of value + 2  $\mu\text{V}$ )

The applied value is millivolts DC derived from BS EN 60584.1 Part 4 1996 Thermocouple Tables.

Calibration procedure MKIP 221 was used.

Standard UKAS certificates used are: DC Volts 29197.

The following results were obtained (See results sheet).

Signed:

Uncertainties are for a confidence probability of not less than 95%.

# M K I S



**Calibration Company**

Milton Keynes Instrumentation Services  
Unit 4 Furtho Court  
Towcester Road  
Old Stratford  
Milton Keynes  
MK19 6AQ

Tel: 01908 568250  
Fax: 01908 564661

## TEST RESULTS

**Certificate No.:** 31811  
**Date of Issue:** 16.04.07

**Customer :** DJB Labcare Limited  
**Instrument Type:** Omega HH506 RA  
**Serial Number:** 05000509

Applied Value	Equivalent Temperature	Indicated Value	Indicated Value
-0.757 mV	-20°C	-20.1°C	-20.1°C
-0.383 mV	-10°C	-10.1°C	-10.1°C
0.00 mV	0°C	-0.2°C	-0.2°C
0.391 mV	10°C	9.8°C	9.8°C
0.790 mV	20°C	19.7°C	19.7°C
2.036 mV	50°C	49.7°C	49.7°C
4.297 mV	100°C	99.7°C	99.7°C
9.288 mV	200°C	199.7°C	199.7°C
14.862 mV	300°C	299.8°C	299.8°C
20.872 mV	400°C	399.8°C	399.8°C

Applied Value	Indicated Value	
	Probe 1	Probe 2
0.0°C	-0.1°C	0.0°C
25°C	24.9°C	25.1°C
37°C	36.8°C	37.1°C
50°C	49.7°C	50.0°C