


# M K I S



**Calibration Company**

Milton Keynes Instrumentation Services  
Unit 4 Furtho Court  
Towcester Road  
Old Stratford  
Milton Keynes  
MK19 6AQ

Tel: 01908 568250  
Fax: 01908 564661

## CERTIFICATE OF CALIBRATION

**CERTIFICATE NO:** 29718

**DATE OF ISSUE:** 11.04.06

<b>Customer:</b>	DJB Labcare Limited	<b>Order Number:</b>	15267
<b>Location:</b>	Newport Pagnell	<b>Serial Number:</b>	H40-1677
<b>Instrument Type:</b>	Seaward IT 1000	<b>Adjustment Required:</b>	No

All measurements were made using equipment whose values are referenced directly or by approved ratiometric procedures to laboratory standards whose values are traceable to National Standards.

The ambient temperature and relative humidity throughout the test was  $21^{\circ}\text{C} \pm 2$  and  $50\% \pm 20$  respectively.

The calibration was carried out in accordance with the general requirements of BS EN ISO/IEC 17025: 2005.

This is to certify that the above listed instrument meets or exceeds as defined by our established calibration procedures, the published manufacturers specification.

The uncertainties refer to the applied and measured values only with no account being taken of the instruments ability to maintain its calibration.

The maximum uncertainties were:

Resistance	
0.1 $\Omega$ to 1.0 $\Omega$	$\pm 0.5\%$
1 $M\Omega$ to 20 $M\Omega$	$\pm 0.05\%$
D C Volts	$\pm 0.05\%$
A C Volts	$\pm 1\%$
A C Current	$\pm 1\%$
Watts	$\pm 1\%$

Applicable UKAS certificates : 020162.

The following results were obtained (see results sheet).

Signed:

Uncertainties are for a confidence probability of not less than 95%.

# M K I S



**Calibration Company**

Milton Keynes Instrumentation Services  
Unit 4 Furtho Court  
Towcester Road  
Old Stratford  
Milton Keynes  
MK19 6AQ

Tel: 01908 568250  
Fax: 01908 564661

## TEST RESULTS

**Certificate No.:** 29718  
**Date of Issue:** 11.04.06

**Customer :** DJB Labcare Limited  
**Instrument Type:** Seaward IT 1000  
**Serial Number:** H40-1677

Earth Bond	Range	Applied Value	Indicated Value
	100 mA	Lead	0.006 $\Omega$
		0.10 $\Omega$	0.108 $\Omega$
		0.20 $\Omega$	0.210 $\Omega$
	4 A	Lead	0.008 $\Omega$
		0.10 $\Omega$	0.109 $\Omega$
		0.20 $\Omega$	0.210 $\Omega$
	6 A	Lead	0.008 $\Omega$
		0.10 $\Omega$	0.108 $\Omega$
		0.20 $\Omega$	0.211 $\Omega$
	12 A	Lead	0.007 $\Omega$
		0.10 $\Omega$	0.108 $\Omega$
		0.20 $\Omega$	0.211 $\Omega$
	20 A	Lead	0.008 $\Omega$
		0.10 $\Omega$	0.109 $\Omega$
		0.20 $\Omega$	0.212 $\Omega$
	25 A	Lead	0.005 $\Omega$
		0.10 $\Omega$	0.107 $\Omega$
		0.20 $\Omega$	0.207 $\Omega$
Insulation		2.00 M $\Omega$	2.05 $\Omega$
		7.00 M $\Omega$	7.34 $\Omega$
		10.0 M $\Omega$	10.3 $\Omega$
Leakage			0.50 mA
			0.48 mA