

CERTIFICATE OF CALIBRATION

ISSUED BY: **M K I S** CALIBRATION COMPANY

DATE OF ISSUE: 29 October 2009


CERTIFICATE NUMBER: 2267



0236

STANDARDS LABORATORY

M K I S



Calibration Company

10 Potters Lane
Kiln Farm

Milton Keynes
MK11 3HE

Tel: 01908 568250
Fax: 01908 564661

Page 1 of 2 pages

Approved
Signatories

R Younger
C Kemp

Signature

Equipment Description:	Tachometer
Manufacturer:	Standard
Type:	ST 6236B
Serial Number:	06019433
Order Number:	17035
Customer:	D J B Labcare
Location:	Newport Pagnell
Date Received:	26 October 2009
Date Calibrated:	28 October 2009

The instrument was kept in the laboratory environment for 24 hours, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $50\% \pm 20\%$ respectively.

The uncertainties reported refer to the applied values only with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

CERTIFICATE OF CALIBRATION

Certificate Number: 2267

UKAS Accredited Calibration Laboratory No. 0236

PAGE 2 OF 2 PAGES

Applied Value	Equivalent Value	Indicated Value
120.06 ms	499.75 RPM	499.8 RPM
60.003 ms	999.95 RPM	1000 RPM
29.995 ms	2000.33 RPM	2000 RPM
20.006 ms	2999.10 RPM	2999 RPM
15.003 ms	3999.20 RPM	3999 RPM
12.004 ms	4998.33 RPM	4999 RPM
10.002 ms	5998.80 RPM	5999 RPM
8.574.8 ms	6997.25 RPM	6997 RPM
7.500 7 ms	7999.25 RPM	8000 RPM
5.999 6 ms	10 000.7 RPM	10 001 RPM
3.000 3 ms	19 998.0 RPM	19 998 RPM
1.499 6 ms	40 010.7 RPM	40 010 RPM
1.000 10 ms	59 994 0 RPM	59 994 RPM

The measurement uncertainties were:

Time $\pm 0.01\% + 1 \text{ LSD}$

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.