

CERTIFICATE OF CALIBRATION

ISSUED BY: **M K I S** CALIBRATION COMPANY

DATE OF ISSUE: 08 March 2010


CERTIFICATE NUMBER: 2301



0236

STANDARDS LABORATORY

M K I S



Calibration Company

10 Potters Lane
Kiln Farm

Milton Keynes
MK11 3HE

Tel: 01908 568250
Fax: 01908 564661

Page 1 of 3 pages

Approved
Signatories

R Younger
C Kemp

Signature

Equipment Description:	Digital Multimeter
Manufacturer:	Uni-T
Type:	UT60-3
Serial Number:	1090518818
Order Number:	Verbal Ref Colin
Customer:	DJB Labcare Limited
Location:	Milton Keynes
Date Received:	05 March 2010
Date Calibrated:	08 March 2010

The instrument was kept in the laboratory environment for 2 days, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $50\% \pm 20\%$ respectively.

The uncertainties reported refer to the applied values only with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

CERTIFICATE OF CALIBRATION

Certificate Number: 2301

UKAS Accredited Calibration Laboratory No. 0236

PAGE 3 OF 3 PAGES

Function	Range	Applied Value	Indicated Value
DC Current	400 μ A	100.00 μ A	100.4 μ A
	4 mA	1.000 0 mA	1.001 mA
	40 mA	10.000 mA	10.05 mA
	400 mA	100.00 mA	100.3 mA
	4 A	1.000 0 A	0.998 A
	10 A	10.000 A	9.96 A

The measurement uncertainties were:

DC Current 100 μ A to 10 A \pm 0.05% + 1 L.S.D.

Function	Range	Applied Value	Indicated Value
AC Current	400 μ A	100.00 μ A	@ 100 Hz 100.5 μ A
	4 mA	1.000 0 mA	@ 100 Hz 1.002 mA
	40 mA	10.000 mA	@ 100 Hz 10.06 mA
	400 mA	100.00 mA	@ 100 Hz 100.3 mA
	4 A	1.000 0 A	@ 100 Hz 0.999 A
	10 A	* 10.000 A	@ 100 Hz * 9.96 A

The measurement uncertainties were:

AC Current 100 μ A to 1 A @ 100 Hz \pm 0.1% + 1 L.S.D.

* Not part of our current accreditation, added for completeness of certificate.

Function	Auto	Applied Value	Indicated Value
Frequency		1.000 Hz	0.999 Hz
		10.000 Hz	9.99 Hz
		100.00 Hz	99.9 Hz
		1.000 0kHz	0.999 kHz
		10.000 kHz	9.99 kHz
		100.00 kHz	99.9 kHz
		1.000 0MHz	0.999 MHz
		10.000 MHz	9.99 MHz

The Measurement uncertainties were:

Frequency 1Hz to 10MHz \pm 0.05% + 1 L.S.D

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

CERTIFICATE OF CALIBRATION

Certificate Number: 2301

UKAS Accredited Calibration Laboratory No. 0236

PAGE 2 OF 3 PAGES

Function	Range	Applied Value	Indicated Value	
DC Volts	400 mV	390.00 mV	390.4 mV	
		-390.00 mV	-390.4 mV	
	4 V	3.900 00 V	3.910 V	
		40 V	10.000 V	10.04 V
	400 V	20.000 V	20.08 V	
		30.000 V	30.12 V	
		39.000 V	39.15 V	
		390.00 V	391.4 V	
		1000 V	1000.0 V	1005 V

The measurement uncertainties were:

DC Volts 390 mV to 1000 V $\pm 0.05\% + 1 \text{ L.S.D.}$

Function	Range	Applied Value	Indicated Value
AC Volts	4 V	3.000 0 V @ 100 Hz	2.994 V
	40 V	30.000 V @ 100 Hz	29.99 V
	400 V	300.00 V @ 100 Hz	299.8 V
	750 V	750.0 V @ 100 Hz	755 V

The measurement uncertainties were:

AC Volts 300 mV to 750 V @ 100 Hz $\pm 0.05\% + 1 \text{ L.S.D.}$

Function	Range	Applied Value	Indicated Value
Resistance	400 Ω	100.00 Ω	99.9 Ω
	4 k Ω	1.000 0 k Ω	0.999 k Ω
	40 k Ω	10.000 k Ω	10.00 k Ω
	400 k Ω	100.00 k Ω	100.1 k Ω
	4 M Ω	1.000 0 M Ω	1.001 M Ω
	40 M Ω	10.000 M Ω	10.01 M Ω

The measurement uncertainties were:

Resistance 100 Ω to 10 M Ω $\pm 0.05\% + 1 \text{ L.S.D.}$