


M K I S



Calibration Company

Milton Keynes Instrumentation Services
10 Potters Lane
Kiln Farm
Milton Keynes
Bucks
MK11 3HE

Tel: 01908 568250
Fax: 01908 564661

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: 39868

DATE OF ISSUE: 06-05-11

Customer:	DJB Labcare Limited	Order Number:	17827
Location:	Newport Pagnell	Serial Number:	None
Instrument Type:	Stop Watch	Adjustment Required:	No

All measurements were made using equipment whose values are referenced directly or by approved radiometric procedures to laboratory standards whose values are traceable to National Standards.

The ambient temperature and relative humidity throughout the test was 20°C ± 2°C and 50% +/- 20 respectively.

The calibration was carried out in accordance with the general requirements of BS EN ISO/IEC 17025: 2005.

This is to certify that the above listed instrument meets or exceeds as defined by our established calibration procedures, the published manufacturers specification.

The uncertainties refer to the applied and measured values only with no account taken of the instrument's ability to maintain its calibration.

The maximum uncertainties were:

Time ± 0.2 s

The frequency was derived from the Anthorn 60 kHz transmission.

The following results were obtained (see results sheet).

Signed:

Uncertainties are for a confidence probability of not less than 95%.

M K I S



Calibration Company

Milton Keynes Instrumentation Services
10 Potters Lane
Kiln Farm
Milton Keynes
Bucks
MK11 3HE

Tel: 01908 568250
Fax: 01908 564661

TEST RESULTS

Certificate No.: 39868 **Customer :** D J B Labcare
Date of Issue: 06-05-11 **Instrument Type:** Stop Watch
Serial Number: None

Indicated Value	Average Value	
10.00 secs	9.99 secs	(Average of 6 readings)
20.00 secs	20.01 secs	
40.00 secs	40.01 secs	
1.00 00 min	59.99 secs	
15.00 00 mins	15.00 00 min	