

CERTIFICATE OF CALIBRATION

ISSUED BY: **M K I S** CALIBRATION COMPANY

DATE OF ISSUE: 08 May 2012

CERTIFICATE NUMBER: 2544



0236

STANDARDS LABORATORY



10 Potters Lane
Kiln Farm

Milton Keynes
MK11 3HE

Tel: 01908 568250
Fax: 01908 564661

Page 1 of 3 pages

Approved
Signatories

R. Younger
C. Kemp

Signature

| | |
|------------------------|---------------------|
| Equipment Description: | Digital Multimeter |
| Manufacturer: | Uni-T |
| Type: | UT60-3 |
| Serial Number: | 1090518818 |
| Order Number: | Ref Paul |
| Customer: | DJB Labcare Limited |
| Location: | Milton Keynes |
| Date Received: | 04 May 2012 |
| Date Calibrated: | 08 May 2012 |

The instrument was kept in the laboratory environment for 2 days, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $50\% \pm 20\%$ respectively.

The uncertainties reported refer to the applied values only with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0236

Certificate Number: 2544

PAGE 2 OF 3 PAGES

| Function | Range | Applied Value | Indicated Value | |
|----------|--------|---------------|-----------------|---------|
| DC Volts | 400 mV | 390.00 mV | 390.2 mV | |
| | | -390.00 mV | -390.2 mV | |
| | 4 V | 3.900 00 V | 3.907 V | |
| | | 10.000 V | 10.02 V | |
| | 40 V | 20.000 V | 20.05 V | |
| | | 30.000 V | 30.08 V | |
| | | 39.000 V | 39.10 V | |
| | | 390.00 V | 391.0 V | |
| | | 400 V | 390.00 V | 391.0 V |
| | | 1000 V | 1000.0 V | 1004 V |

The measurement uncertainties were:

DC Volts 390 mV to 1000 V $\pm 0.05\% + 1 \text{ L.S.D.}$

| Function | Range | Applied Value | Indicated Value | |
|----------|-------|--------------------|-----------------|-------|
| AC Volts | 4 V | 3.000 0 V @ 100 Hz | 2.991 V | |
| | 40 V | 30.000 V @ 100 Hz | 29.95 V | |
| | 400 V | 300.00 V @ 100 Hz | 299.4 V | |
| | 750 V | 750.0 V @ 100 Hz | | 753 V |

The measurement uncertainties were:

AC Volts 300 mV to 750 V @ 100 Hz $\pm 0.05\% + 1 \text{ L.S.D.}$

| Function | Range | Applied Value | Indicated Value |
|------------|----------------|--------------------|------------------|
| Resistance | 400 Ω | 100.00 Ω | 99.8 Ω |
| | 4 k Ω | 1.000 0 k Ω | 0.999 k Ω |
| | 40 k Ω | 10.000 k Ω | 10.00 k Ω |
| | 400 k Ω | 100.00 k Ω | 100.0 k Ω |
| | 4 M Ω | 1.000 0 M Ω | 1.001 M Ω |
| | 40 M Ω | 10.000 M Ω | 10.02 M Ω |

The measurement uncertainties were:

Resistance 100 Ω to 10 M Ω $\pm 0.05\% + 1 \text{ L.S.D.}$

CERTIFICATE OF CALIBRATION

Certificate Number: 2544

UKAS Accredited Calibration Laboratory No. 0236

PAGE 3 OF 3 PAGES

| Function | Range | Applied Value | Indicated Value |
|------------|---------------|----------------|-----------------|
| DC Current | 400 μ A | 100.00 μ A | 100.3 μ A |
| | 4 000 μ A | 1000.0 μ A | 1001 μ A |
| | 40 mA | 10.000 mA | 10.05 mA |
| | 400 mA | 100.00 mA | 100.3 mA |
| | 4 A | 1.000 0 A | 0.997 A |
| | 10 A | 10.000 A | 9.96 A |

The measurement uncertainties were:

DC Current 100 μ A to 10 A $\pm 0.05\% + 1$ L.S.D.

| Function | Range | Applied Value | Indicated Value |
|------------|--------------|----------------|------------------------|
| AC Current | 400 μ A | 100.00 μ A | @ 100 Hz 100.4 μ A |
| | 4000 μ A | 1000.0 μ A | @ 100 Hz 1001 μ A |
| | 40 mA | 10.000 mA | @ 100 Hz 10.06 mA |
| | 400 mA | 100.00 mA | @ 100 Hz 100.3 mA |
| | 4 A | 1.000 0 A | @ 100 Hz 0.998 A |
| | 10 A | * 10.000 A | @ 100 Hz * 9.95 A |

The measurement uncertainties were:

AC Current 100 μ A to 1 A @ 100 Hz $\pm 0.1\% + 1$ L.S.D.

* Not part of our current accreditation, added for completeness of certificate.

| Function | Auto | Applied Value | Indicated Value |
|-----------|------|---------------|-----------------|
| Frequency | | 1.000 Hz | 0.999 Hz |
| | | 10.000 Hz | 9.99 Hz |
| | | 100.00 Hz | 99.9 Hz |
| | | 1.000 kHz | 0.999 kHz |
| | | 10.000 kHz | 9.99 kHz |
| | | 100.00 kHz | 99.9 kHz |
| | | 1.000 MHz | 0.999 MHz |
| | | 10.000 MHz | 9.99 MHz |
| | | | |

The Measurement uncertainties were:

Frequency 1Hz to 10MHz $\pm 0.05\% + 1$ L.S.D

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.