CERTIFICATE OF CALIBRATION

Date Of Issue 16th December 2013

Certificate Number 161213/N15



0601

Certificate Issued By:



Electronic Temperature Instruments Ltd Easting Close, Worthing West Sussex BN14 8HQ England Tel Fa

lephone +44 01903 202151
acsimile +44 01903 202445
email: sales@etiltd.co.uk
Website: www.etiltd.co.uk

Page 1 of 1 pages	
Approved Signatory	Signature
J. Carswell	18 Curusal
R. Collins	0
G. Hills	

Customer Name: DJB LABCARE LIMITED

Address:

UNIT 20 HOWARD WAY INTERCHANGE PARK **NEWPORT PAGNELL** BUCKINGHAMSHIRE

MK16 9QS

Order Number:

19225

Ref Number: 384/38297

Date Received:

13th December 2013

Date Calibrated:

16th December 2013

Ambient Temperature: 22 °C ± 2 °C

Ambient Humidity: < 60 %rh

Temperature Scale:

International Temperature Scale of 1990

Inst Description: AMIPROBE TMD-56 INDICATOR

Sensor Type:

2000mm TYPE "T" EXPOSED JUNCTION WIRE PROBE

Inst Serial Number: 13040325

Sensor Serial Number: P325

Calibrated Range/Scale: -50 °C to + 100 °C / 0.1 °C RESOLUTION

Procedure: The Instrument was stabilised at ambient temperature, then calibrated by comparison

with two traceable references in stirred liquid baths. Immersion depth

Results: Test Temperature °C

Instrument Reading °C

- 49.0	- 50.00
0.4	0.00
37.1	37.00
100.3	100.00

Uncertainty of Measurement ± 0.15 °C + Instrument Resolution.

End of Report

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability or long term stability

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurements to recognised national standards, and to units of measurement realised at the National Physical Laboratory and other recognised national standards laboratories, including the National Institute of Standards and Technology (NIST).

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.