

CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

DATE OF ISSUE: 27th August 2025

CERTIFICATE No: 937672



Lambda
CALIBRATION LTD

11-13 Chorley Central
Business Park
Stump Lane
Chorley
PR6 0BL
Tel: 01257 244 670

Page 1 of 2

APPROVED SIGNATORY

C Reed E Santos R Armitage
K Quigley D Pilkington

Customer: DJB Labcare Ltd, Newport Pagnell, MK16 9QS
Item No: 539276
Description: Frequency Counter
Model/Range: PFM3000
Manufacturer: Aim TTi
Date of Cal: 27/08/2025
Basis: E-900
Equipment Used: GPS Frequency Reference (LGPS), Function Generator (LFG-02, LSG-03)
Temp/Humidity: 20°C ± 2°C, <80%rh

Initial Inspection on Receipt: (For information only – not part of our UKAS accredited schedule).

Unit Under Test (UUT) Condition	Functional.
Condition of Supplied Leads	None Supplied.
Battery	Replaced.

Summary of Results:

Pre Calibration Status	The reported results fall within the specified tolerances
Post Calibration Status	The reported results fall within the specified tolerances
Adjustments	No
Repairs	No
Test Leads Used	Laboratory test leads were used in the calibration.
Comments	-

Measured results and measurement uncertainties are detailed on the following pages.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. Unless otherwise stated: [1] The 'Compliance Statement' is based on 'simple acceptance' (result vs tolerance) with the relevant calibration uncertainty being no greater than the tolerance. [2] Reported activities were carried out at the address detailed in the header. [3] The results relate only to the items calibrated. This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and / or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

UKAS ACCREDITED CALIBRATION LABORATORY No: 0495

CERTIFICATE No:

937672

Page 2 of 2

Frequency Accuracy (Input A)

UUT Range	UUT Gate	Applied Frequency (Hz)	UUT Indication (Hz)
Freq A	10s	10.0000000	10.0000081
Freq A	10s	100.000000	100.000086
Freq A	10s	1000.0000	1000.00090
Freq A	10s	1.000000M	1000.00092k
Freq A	0.3s	10.000000M	10.00001M
Freq A	1s	10.000000M	10.000009M
Freq A	10s	10.000000M	10.0000093M
Freq A	10s	100.00000M	100.000094M

Frequency Accuracy (Input B)

UUT Range	UUT Gate	Applied Frequency (MHz)	UUT Indication (MHz)
Freq B	10s	80.000000	80.000078
Freq B	10s	500.000000	500.00050
Freq B	0.3s	1000.00000	1000.001
Freq B	1s	1000.00000	1000.0010
Freq B	10s	1000.00000	1000.00105
Freq B	10s	2000.00000	2000.0022
Freq B	10s	2700.00000	2700.0029

Period Accuracy (Input A)

UUT Range	UUT Gate	Applied Time Period (s)	UUT Indication (s)
Period A	0.3s	100.00000n	99.9999n
Period A	1s	100.000000n	99.99989n
Period A	10s	100.0000000n	99.999890n

Manufacturers Tolerance applicable to above results: $\pm(1 \text{ digit} + \text{Timebase Accuracy (7ppm)})$

End of Results

Estimated Uncertainty of Measurement:

Frequency: $\pm(0.29\text{ppm} + 1.5\mu\text{Hz} + 2 \text{ digits})$