



Calibration Laboratory and Location of Calibration

Certificate number: 013-151025

DJB Labcare Ltd 12 Howard Way Cromwell Business Centre Newport Pagnell Bucks MK16 9QS

> Tel: 01908 612598 Fax: 01908 217974

E-mail: service@djblabcare.co.uk Website: www.djblabcare.co.uk

Certificate of Calibration

Equipment: Fastime 3 Stopwatch

Serial number: 013

The results refer only to the equipment listed above

Date of Issue: 15/10/2025 Date of Calibration: 15/10/2025

Ambient Temperature: 19.7°C

Results:

Set Value	Measured Value
05:00.00	04:59.96
10:00.00	09:59.93
60:00.00	60:00.12

Method:

A time interval standard is used and compared to the DUT's display, carried out at 5,10 and 60 min intervals.

Uncertainty: 0.48s

Uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Test Equipment Details:

BT Speaking Clock

Approved by: C Bradnam

END OF RESULTS

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.