

# CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

DATE OF ISSUE: 12 December 2025 CERTIFICATE NUMBER: **1190881A**



0654



Unit 5 Octavian Way  
Team Valley Trading Estate  
Gateshead  
NE11 0HZ

Tel: +44 (0)191 4875951

Page 1 of 3

Approved Signatory

Electronically Authorised Document

P K CLARK     R LOMAX  
 R J WADE     J FRYER  
 M A FROST     M FOY  
 M YEATES

## CUSTOMER

DJB LABCARE LTD  
UNIT 12 HOWARD WAY  
CROMWELL BUSINESS CENTRE  
NEWPORT PAGNELL  
BUCKINGHAMSHIRE  
MK16 9QS  
UNITED KINGDOM

## Manufacturer

### Description

### Model

### Serial No.

### Ident No.

### Date Received

### Calibrated By

### Date Of Calibration

### Order No.

STANDARD

TACHOMETER OPTICAL

AT-6

200552135

UNKNOWN

12 NOVEMBER 2025

MALCOLM FOY

17 NOVEMBER 2025

NS24388

## INSTRUMENT CONDITION

Adjustments Made                      No  
Repairs Made                              No

## ENVIRONMENT

The instrument was placed in the Laboratory environment for a minimum period of 4 hours prior to calibration.

The ambient conditions were: 18 to 23°C and 45 %rh  $\pm$  15 %rh.

## STABILITY

The results contained in this Certificate refer to the measurements made at the time of test and not to the instrument's ability to maintain calibration.

## PROCEDURE

Measurements were performed in accordance with CMR Ltd's calibration procedure No. 1053

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

## Supplement to Calibration Certificate

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654



CERTIFICATE NUMBER

**1190881A**

**Page 2 of 3**

## Calibration Equipment Used:

<b>Cert Number</b>	<b>Ident Number</b>	<b>Model</b>	<b>Serial Number</b>	<b>Test Equipment Calibration Due</b>
1187001IH	143	33220A	MY44025316	29 Sep 2026
1149140IH	475	A002206.G1	13572091	3 Sep 2026

## Notes:

Results relate only to the items calibrated.

This certificate has been re-issued. The environmental conditions statement was missing.

Missing statement: " The ambient conditions were: 18 to 23°C and 45 %rh ± 15 %rh"

# CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654



CERTIFICATE NUMBER

**1190881A**

## RESULT SHEET 1053-NC-UKAS-44 : STANDARD AT-6 OPTICAL TACHOMETER

### AS FOUND

**BATTERY REPLACED: NO**

The Revolution Per Minute (RPM) was simulated using a precision Frequency Generator with an attached optical light source.

### **RPM**

Range	Simulated	Indicated	Units	Indicated Deviation	Measurement Uncertainty ±
2 to 99999RPM	600.000	600.0	RPM	0.000	0.064
	1000.00	1000	RPM	0.00	0.58
	2500.00	2500	RPM	0.00	0.58
	5000.00	5000	RPM	0.00	0.58
	7499.99	7500	RPM	0.01	0.58
	9999.79	10000	RPM	0.21	0.58
	14999.46	15000	RPM	0.54	0.58
	19999.13	20000	RPM	0.87	0.58
	24998.75	25000	RPM	1.25	0.58
	29998.42	30000	RPM	1.58	0.58

Manufacturer stated Accuracy:  $\pm(0.05\%+1 \text{ digit})$

**-END OF RESULTS-**