## CERTIFICATE OF CALIBRATION

Date Of Issue 11th July 2014 Certificate Number 110714/N1



0601

Certificate Issued By:



Electronic Temperature Instruments Ltd Easting Close, Worthing West Sussex BN14 8HQ England Telephone +44 01903 202151 Facsimile +44 01903 202445 email: sales@etiltd.co.uk Website: www.etiltd.co.uk

Signature
,
Stills

Customer Name: DJB LABCARE

Address:

20 HOWARD WAY

CROMWELL BUSINESS CENTRE

INTERCHANGE PARK

NEWPORT PAGNELL, BUCKINGHAMSHIRE

MK16 9QS

Order Number:

Jude Knapman

Ref Number: 414/41218

Date Received:

4th July 2014

Date Calibrated:

11th July 2014

Ambient Temperature:

22 °C ± 2 °C

Ambient Humidity:

< 60 %rh

Temperature Scale:

International Temperature Scale of 1990

Inst Description:

AMPROBE TMD56 INDICATOR

Sensor Type:

1000mm TYPE 'K' EXPOSED JUNCTION WIRE PROBE

Inst Serial Number:

13110363

Sensor Serial Number: P0363

Calibrated Range/Scale:

- 50 °C to + 100 °C / 0.1 °C RESOLUTION

Procedure:

The Instrument was stabilised at ambient temperature, then calibrated by comparison with two traceable references in stirred liquid baths. Immersion depth

Results:

Test Temperature °C

Instrument Reading °C

- 49.3	- 50.00
0.4	0.00
37.2	37.00
75.0	75.00
99.4	100.00

Uncertainty of Measurement ± 0.15 °C + Instrument Resolution.

End of Report

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.

The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability or long term stability.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurements to recognised national standards, and to units of measurement realised at the National Physical Laboratory and other recognised national standards laboratories, including the National Institute of Standards and Technology (NIST).

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.