



8807



**Calibration Laboratory and Location of Calibration**

DJB Labcare Ltd  
12 Howard Way  
Cromwell Business Centre  
Newport Pagnell  
Bucks MK16 9QS

Tel: 01908 612598  
Fax: 01908 217974

E-mail: [service@djblabcare.co.uk](mailto:service@djblabcare.co.uk)  
Website: [www.djblabcare.co.uk](http://www.djblabcare.co.uk)

# Certificate of Calibration

Certificate number: 016-080626

Equipment: Fastime 3 Stopwatch  
Serial number: 016  
The results refer only to the equipment listed above

Date of Issue: 08/06/2026  
Date of Calibration: 08/06/2026

Ambient Temperature: 22.1°C

**Results:**

Set Value	Measured Value
05:00.00	04:59.94
10:00.00	10:00.06
60:00.00	60:00.09

**Method:**

A time interval standard is used and compared to the DUT's display, carried out at 5, 10 and 60 min intervals.

**Uncertainty: 0.48s**

Uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

**Test Equipment Details:**

BT Speaking Clock

Approved by: C Bradnam

**END OF RESULTS**

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.