

Certificate of Calibration

Certificate number: C1505-290Equipment: RS 278-682 Stopwatch
Serial number: 002Date: 01/05/2015
Calibration frequency: Every 12 Months

Tested by: Martin Paragreen

Results:

Set Value	Measured Value
5:00.00	04:59.86
10:00.00	09:59.75
60:00.00	59:59.79

Method of test:

All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith

