

Certificate of Calibration

Certificate number: **C1505-287**Equipment: TIM 901S Stopwatch
Serial number: 008Date: 01/05/2015
Calibration frequency: Every 12 months

Calibrated by: Martin Paragreen

Results:

Set Value	Measured Value
05:00.00	05:00.30
10:00.00	10:00.06
60:00.00	60:00.15

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith

