

CERTIFICATE OF CALIBRATION

Issued By Combined Precision Components Plc

Date of Issue 09 October 2007

Certificate Number
CPC34938

Page 1 of 2 Pages



NPL HILLHOUSE IBC
PO BOX 4, STANDARDS BUILDING
THORNTON CLEVELEYS
LANCASHIRE, FY5 4QD
TEL :- 01253 820200 FAX :- 01253 827700

Approved Signatory

A.Foster L.Willis P.Stevens D.Berkley N.Harrison

Customer : DJB1217 DJB LAB CARE
20 HOWARD WAY, INTERCHANGE PARK
NEWPORT PAGNALL, BUCKS MK16 9QS

Date Received : 09 October 2007

Instrument -	System ID :	S4B06019415	Job Number :	0005064
	Description :	Tachometer		
	Manufacturer :	Standard		
	Model Number :	ST6234B		
	Serial Number :	06019415		
	Procedure Version :	GG3		

Environmental Conditions

Temperature : 20°C +/- 3°C
Relative Humidity : 50% +/- 10%

Mains Voltage : 240V +/- 10V
Mains Frequency : 50Hz +/- 1Hz

Comments

Instrument was placed in lab and allowed to stabilise before calibration.

All Tests Passed calibration.

Traceability Information

<i>Instrument description</i>	<i>Serial number</i>	<i>Certificate number</i>	<i>Cal. Date</i>	<i>Cal. Period</i>
5500 Multifunction Calibrator	718509	11092	26/02/2007	52

Calibrated By : N.Harrison

Date of Calibration : 09 October 2007

This certificate provides traceability of measurement to recognised National Standards, and to the units of measurement realised at the National Physical Laboratory or other recognised National Standards laboratories.

Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior written approval of the issuing laboratory. This certificate complies with the requirements of B.S.5781, part 1 1992. (ISO 10012 - 1 1992).

CERTIFICATE OF CALIBRATION

Certificate Number
CPC34938

Page 2 of 2 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
Display Ok	---	---	Pass	
RPM/TOT Switch Ok	---	---	Pass	
Measure Button Ok	---	---	Pass	
MEM Button Ok	---	---	Pass	
Opto-Coupler Tests				
RPM	1.5rpm	1 000rpm	1 000rpm	Pass
..	2rpm	2 000rpm	2 000rpm	Pass
..	2.5rpm	3 000rpm	2 999rpm	Pass
..	3rpm	4 000rpm	4 000rpm	Pass
..	3.5rpm	5 000rpm	5 000rpm	Pass
..	4rpm	6 000rpm	6 000rpm	Pass
..	4.5rpm	7 002rpm	7 002rpm	Pass
..	5rpm	7 998rpm	7 997rpm	Pass
..	5.5rpm	9 000rpm	8 999rpm	Pass
..	6rpm	10 002rpm	10 002rpm	Pass
..	8.5rpm	15 000rpm	15 001rpm	Pass
..	11rpm	19 998rpm	19 998rpm	Pass
..	13.5rpm	25 002rpm	25 000rpm	Pass
..	16rpm	30 000rpm	30 003rpm	Pass
..	21rpm	40 002rpm	40 005rpm	Pass
..	31rpm	60 000rpm	60 000rpm	Pass
..	46rpm	90 000rpm	89 998rpm	Pass

Uncertainties

RPM : 0.001%