CERTIFICATE OF CALIBRATION

ISSUED BY: MKIS CALIBRATION COMPANY

DATE OF ISSUE: 02 June 2016

CERTIFICATE NUMBER: 2915



0236

STANDARDS LABORATORY

MKIS Calibration Company 10 Potters Lane Kiln Farm

Milton Keynes MK11 3HE

Tel: 01908 568250 Fax: 01908 564661 Page 1 of 2 pages

Approved Signatories R Younger C.Kemp

Signature

Equipment Description: Optical Tachometer

Manufacturer:

AT-6 Type:

Serial Number:

12078449

Order Number:

NS 20633

Customer: Location:

Newport Pagnell

Standard

Date Received:

25 May 2016

D J B Labcare

Date Calibrated:

01 June 2016

The instrument was kept in the laboratory environment for 2 days, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was 20°C ± 2°C and $50\% \pm 20\%$ respectively.

The uncertainties reported refer to the applied and indicated values only with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

CERTIFICATE OF CALIBRATION

Certificate Number: 2915

UKAS Accredited Calibration Laboratory No. 0236

PAGE 2 OF 2 PAGES

Applied Val	lue Eq	uivalent `	Value	Indicated `	Value
99.999	ms	600.00	RPM	600.0	RPM
59.999	ms	1000.01	RPM	1000	RPM
24.002	ms	2499.8	RPM	2500	RPM
12.000	ms	5000.0	RPM	5000	RPM
8.000 2	ms	7499.8	RPM	7500	RPM
5.999 9	ms	10000.2	RPM	10 001	RPM
4.285 66	ms	14000.2	RPM	14 001	RPM
2.999 93	ms	20000.5	RPM	20 001	RPM
2.399 99	ms	25000.1	RPM	25 001	RPM
1.999 99	ms	30000.1	RPM	30 001	RPM
1.500 01	ms	39999.7	RPM	40 000	RPM
1.200 00	ms	50000.0	RPM	50 001	RPM
1.000 02	ms	59998.2	RPM	59 999	RPM

The measurement uncertainties were:

Time

 \pm 0.1 RPM + 1 LSD

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.