

CERTIFICATE OF CALIBRATION

ISSUED BY: **M K I S** CALIBRATION COMPANY

DATE OF ISSUE: 22 October 2008


CERTIFICATE NUMBER: 2137



0236

STANDARDS LABORATORY

M K I S



Calibration Company

10 Potters Lane

Kiln Farm
Milton Keynes
Mk11 3HE

Tel: 01908 568250
Fax: 01908 564661

Page 1 of 2 pages

Approved
Signatories

R Younger
C Kemp

Signature

Equipment Description:	Digital Thermometer
Manufacturer:	Omega Engineering
Type:	HH 506RA
Serial Number:	05001201
Order Number:	16504
Customer:	DJB Labcare Limited
Location:	Newport Pagnell
Date Received:	12 September 2008
Date Calibrated:	22 October 2008

The instrument was kept in the laboratory environment for 24 hours, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $50\% \pm 10\%$ respectively.

The uncertainties reported refer to the applied values only with no account being taken of the instruments ability to maintain its calibration.

The applied value is millivolts DC derived from BS EN 60584.1 Part 5 1996 Thermocouple Tables.

Remarks: No adjustments were made.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0236

Certificate Number: 2137

PAGE 2 OF 2 PAGES

Applied Value	Equivalent Temperature	Indicated Value	
		T 1	T 2
-0.757 mV	-20°C	-20.2°C	-20.2°C
-0.388 mV	-10°C	-10.2°C	-10.2°C
0.00 mV	0°C	-0.1°C	-0.1°C
0.391 mV	10°C	10.0°C	10.0°C
0.790 mV	20°C	19.8°C	19.8°C
1.196 mV	30°C	29.8°C	29.8°C
1.486 mV	37°C	36.8°C	36.8°C
1.612 mV	40°C	39.9°C	39.9°C
2.036 mV	50°C	49.7°C	49.8°C
4.279 mV	100°C	99.9°C	99.9°C
6.704 mV	150°C	149.9°C	150.0°C
9.288 mV	200°C	199.9°C	199.9°C
12.013 mV	250°C	249.9°C	249.9°C
14.862 mV	300°C	299.9°C	299.9°C

The maximum uncertainties were:

-20°C to 300°C $\pm 0.2^\circ\text{C} + 0.5 \text{ L.S.D.}$

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Supplement to UKAS report No 2137

Date 22 October 2008

Page 1 of 1

The following parameters are not included in our current UKAS Accreditation and are not part of Report: 2137

Applied Value		Indicated Value	
		T 1	T 2
T/C Supplied Type T	0°C	0.0°C	0.0°C
	37.0°C	36.9°C	36.9°C