



Certificate of Calibration

Certificate number: 110818

Equipment: Marble Stopwatch
Serial number: 011

Date: 22/08/18
Calibration frequency: Every 12 months

Calibrated by: Caitriona Hurd

Results:

Set Value	Measured Value
05:00.00	05:00.12
10:00.00	10:00.04
60:00.00	59:59.89

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam