



Certificate of Calibration

Certificate number: 190818

Equipment: Fastime Stopwatch
Serial number: 019

Date: 28/08/18
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

Results:

Set Value	Measured Value
05:00.00	05:00.04
10:00.00	10:00.04
60:00.00	59:59.92

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam