



20 Howard Way  
Cromwell Business Centre  
Interchange Park  
Newport Pagnell  
Bucks MK16 9QS

Tel: 01908 612598  
Fax: 01908 217974

E-mail: [service@djblabcare.co.uk](mailto:service@djblabcare.co.uk)  
Website: [www.djblabcare.co.uk](http://www.djblabcare.co.uk)

## Certificate of Calibration

Certificate number: DJB203-1

**Equipment:** Samson SSA-20 CO2 Analyser  
**Serial number:** 161009

Analysers are calibrated on the morning of any pending CO2 calibrations in accordance with DJB Labcare procedures

### Method

All calibrations are carried out in a temperature stable laboratory using reference gases providing traceability to UKAS ISO17025 standards and in accordance with the manufacturers procedures

Certified traceable sample gas from EfecTech, Laboratory Number 0590 Co2 5% - Cylinder number D690633. Certificate Number: 18/0796/01

Date of calibration 12<sup>th</sup> June 2018.

Approved by: Colin Bradnam

A handwritten signature in purple ink, appearing to read 'Colin Bradnam', written over a faint dotted grid.