



Certificate of Calibration

Certificate number: 130818

Equipment: Fastime 3 Stopwatch
Serial number: 013

Date: 17/08/18
Calibration frequency: Every 12 months

Calibrated by: Caitriona Hurd

Results:

Set Value	Measured Value
05:00.00	05:00.02
10:00.00	10:00.08
60:00.00	60:00.16

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam