



Certificate of Calibration

Certificate number: 121018

Equipment: Fastime 3 Stopwatch
Serial number: 012

Date: 17/10/18
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

Results:

Set Value	Measured Value
05:00.00	05:00.15
10:00.00	09:59.93
60:00.00	59:59.96

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam