CERTIFICATE OF CALIBRATION

ISSUED BY: MKIS CALIBRATION COMPANY

DATE OF ISSUE: 18 October 2018

CERTIFICATE NUMBER: 3110



STANDARDS LABORATORY



10 Potters Lane Kiln Farm

Milton Keynes MK11 3HE

Tel: 01908 568250 Fax: 01908 564661

Page 1 of	pages		
Approved Signatories	1	A.J.Cox R Younger	
Signature	My	CKemp	

Optical Tachometer **Equipment Description:**

Standard Manufacturer: **AT-6** Type:

12078435 Serial Number: Order Number: NS 21935 DJB Labcare Customer:

Newport Pagnall Location: 16 October 2018 Date Received: 18 October 2018 Date Calibrated:

The instrument was kept in the laboratory environment for 24 hours, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was 20°C ± 2°C and $50\% \pm 20\%$ respectively.

The uncertainties reported refer to the applied and obtained values only, with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

MKIS UKAS Calibration Procedures Used: 5 (5.6)

CERTIFICATE OF CALIBRATION

Certificate Number: 3110

UKAS Accredited Calibration Laboratory No. 0236

PAGE 2 OF 2 PAGES

Applied Value	Equivalent	Value	Indicated '	Value
99.993 ms	600.04	RPM	600.0	RPM
59.997 ms	1000.05	RPM	1000	RPM
24.005 ms	2499.5	RPM	2499	RPM
12.000 3 ms	4998.7	RPM	4999	RPM
8.000 08 ms	7499.9	RPM	7500	RPM
5.999 95 ms	10 000.1	RPM	10 000	RPM
4.000 04 ms	14 999.8	RPM	15 000	RPM
3.000 06 ms	19 999.6	RPM	20 000	RPM
2.400 07 ms	24 999.3	RPM	24 999	RPM
2.000 08 ms	29 998.8	RPM	29 999	RPM

The measurement uncertainties were:

Time

 \pm 0.1 RPM + 1 LSD

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.