



Certificate of Calibration

Certificate number: 14/7219

Equipment: Fastime 3 Stopwatch
Serial number: 014

Date: 07/02/19
Calibration frequency: Every 12 months

Calibrated by: Jude Huckle

Results:

Set Value	Measured Value
05:00.00	04:59.96
10:00.00	09:59.93
60:00.00	59:59.99

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam