



Certificate of Calibration

Certificate number: 10/300623

Equipment: Fastime 3 Stopwatch
Serial number: 010

Date: 30/06/2023
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

Results:

Set Value	Measured Value
05:00.00	05:00.13
10:00.00	10:00.13
60:00.00	60:00.01

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam

