

# M K I S



Calibration Company

Milton Keynes Instrumentation Services  
10 Potters Lane  
Kiln Farm  
Milton Keynes  
Bucks  
MK11 3HE

Tel: 01908 568250  
Fax: 01908 564661

## CERTIFICATE OF CALIBRATION

**CERTIFICATE NO:** 46310

**DATE OF ISSUE:** 02-06-14

**Customer:** DJB Labcare Limited

**Order Number:** 19509

**Location:** Newport Pagnell

**Serial Number:** 40462 08/09

**Instrument Type:** TIM 901R Stop Watch

**Adjustment Required:** No

All measurements were made using equipment whose values are referenced directly or by approved radiometric procedures to laboratory standards whose values are traceable to National Standards.

The ambient temperature and relative humidity throughout the test was  $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$  and 50% +/- 20 respectively.

The calibration was carried out in accordance with the general requirements of BS EN ISO/IEC 17025: 2005.

This is to certify that the above listed instrument meets or exceeds as defined by our established calibration procedures, the published manufacturers specification.

The uncertainties refer to the applied and measured values only with no account taken of the instrument's ability to maintain its calibration.

The maximum uncertainties were:

Time  $\pm 0.02$  s

The frequency was derived from the Anthorn 60 kHz transmission.

The following results were obtained (see results sheet).

Signed:

Uncertainties are for a confidence probability of not less than 95%.

# M K I S



**Calibration Company**

Milton Keynes Instrumentation Services  
10 Potters Lane  
Kiln Farm  
Milton Keynes  
Bucks  
MK11 3HE

Tel: 01908 568250  
Fax: 01908 564661

## TEST RESULTS

**Certificate No.:** 46310      **Customer :** DJB Labcare Limited  
**Date of Issue:** 02-06-14      **Instrument Type:** TIM 901R Stop Watch  
**Serial Number:** 40462 08/09

Indicated Value		Measured Value			Average of Average Values
		Set 1	Set 2	Set 3	
1.00 mins	Average Value	1m 00.02 s	1m 00.01 s	0m 59.99 s	<b>1m 00.01 s</b>
5.00 mins	Average Value	5m 00.01 s	5m 00.02 s	5m 00.01 s	<b>5m 00.01 s</b>
10.00 mins	Average Value	10m 00.01 s	10m 00.01 s	10m 00.02 s	<b>10m 00.01 s</b>
20.00 mins	Average Value	19m 59.98 s	20m 00.02 s	20m 00.02 s	<b>20m 00.01 s</b>
30.00 mins	Average Value	30m 00.01 s	29m 59.99 s	30m 00.03 s	<b>30m 00.01 s</b>