

# CERTIFICATE OF CALIBRATION

ISSUED BY: **M K I S** CALIBRATION COMPANY

DATE OF ISSUE: 10 March 2015


CERTIFICATE NUMBER: 2806



0236

## STANDARDS LABORATORY

# MKIS



Calibration Company

10 Potters Lane  
Kiln Farm

Milton Keynes  
MK11 3HE

Tel: 01908 568250  
Fax: 01908 564661

Page 1 of 2 pages

Approved  
Signatories

R Younger

C Kemp

Signature

Equipment Description:	Digital Tachometer
Manufacturer:	Standard
Type:	AT-6
Serial Number:	12078389
Order Number:	20006
Customer:	D J B Labcare
Location:	Newport Pagnell
Date Received:	09 March 2015
Date Calibrated:	10 March 2015

The instrument was kept in the laboratory environment for 24 hours, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was  $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$  and  $50\% \pm 20\%$  respectively.

The uncertainties reported refer to the applied and indicated values only with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

# CERTIFICATE OF CALIBRATION

Certificate Number: 2806

UKAS Accredited Calibration Laboratory No. 0236

PAGE 2 OF 2 PAGES

Applied Value	Equivalent Value	Indicated Value
100.004 ms	599.97 rpm	600.0 rpm
59.998 ms	1000.03 rpm	1000 rpm
30.002 ms	1999.9 rpm	2000 rpm
14.998 ms	4000.5 rpm	4001 rpm
10.000 5 ms	5999.7 rpm	6000 rpm
6.000 12 ms	9 999.8 rpm	10 000 rpm
3.999 88 ms	15 000.5 rpm	15 000 rpm
2.999 90 ms	20 000.7 rpm	20 001 rpm
1.500 04 ms	39 998.9 rpm	39 999 rpm

The measurement uncertainties were:

RPM  $\pm 0.05$  RPM + 1 LSD

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.