



## Certificate of Calibration

Certificate number: 110316

Equipment: Fastime 3 Stopwatch  
Serial number: 011

Date: 11/03/2016  
Calibration frequency: Every 12 months

Calibrated by: Martin Paragreen

### Results:

Set Value	Measured Value
05:00.00	04:59.87
10:00.00	10:00.04
1:00:00.00	59:59.89

Method of test:  
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith

A handwritten signature in black ink, appearing to read 'P Smith'.