

Certificate of Calibration

Certificate number: DJB207**Equipment:** G100 Gas Analyser**Serial number:** INO5425

Analysers are calibrated on the morning of any pending CO2 calibrations in accordance with DJB Labcare procedures

Method

All calibrations are carried out in a temperature stable laboratory using reference gases providing traceability to UKAS ISO17025 standards and in accordance with the manufacturers procedures

Certified traceable sample gas from EffecTech, Laboratory Number 0590

Co2 5% - Cylinder number D591004, Date of calibration 19th July 2017, Date of expiry 18th July 2022

Approved by: P Smith

