

CERTIFICATE OF CALIBRATION

Issued By Pullman Instruments (UK) Ltd.

Certificate Number 5751

Date of Issue 01 August 2019



0683

Approved Signatory

Page 1 of 2 Pages



Pullman Instruments (UK) Ltd.

Chatsworth House,

Chatsworth Terrace,

Harrogate, HG1 5HT

Tel 01423 720360 Email info@pullman.co.uk

A.J.Cox

M. Conboy

Customer : DJB LABCARE
20 HOWARD WAY, CROMWELL BUSINESS CENTRE
INTERCHANGE PARK

Date Received : 31 July 2019

Instrument :	System ID :	12078449	Job Number :	456495-1
	Description :	Digital Tachometer	Ref. Number :	4421-1
	Manufacturer :	STANDARD	Site :	
	Model Number :	AT-6	Location :	
	Serial Number :	12078449		
	Procedure Version :	20714v01/N		

Environmental Conditions

Temperature : 20°C +/- 2°C
Relative Humidity : 50% +/- 15%

Mains Voltage : 230V +/- 20V
Mains Frequency : 50Hz +/- 1Hz

Comments

Instrument was allowed to stabilise for at least 12 hours prior to calibration.

Measurements traceable to national standards.

Calibration Information

The instrument was calibrated against laboratory standards whose values are traceable to recognised National Standards. The uncertainty limits quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Calibrated By : M.Conboy

Date of Calibration : 01 August 2019

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0683
AS FOUND RESULTS

Certificate Number
5751

Page 2 of 2 Pages

Test Title	Applied Value	Reading	Uncertainties
RPM			
500	500.0rpm	500.0rpm	±0.5 RPM
1000	1 000.0rpm	1 000.0rpm	±0.5 RPM
2500	2 500.0rpm	2 500.0rpm	±0.5 RPM
5000	5 000.0rpm	5 000.0rpm	±0.5 RPM
7500	7 500.0rpm	7 500.0rpm	±0.5 RPM
10000	10 000rpm	10 000rpm	±0.5 RPM
15000	15 000rpm	15 000rpm	±0.5 RPM
20000	20 000rpm	20 000rpm	±0.5 RPM

End Of Calibration.