

Unit 12
Howard Way
Cromwell Business Centre
Interchange Park
Newport Pagnell
Bucks
MK16 9QS
Tel: 01908 612598
Fax: 01908 217974
E-mail: service@djblabcare.co.uk
Website: www.djblabcare.co.uk

Certificate of Calibration

Certificate number: DJB040919-03

Equipment: G100 Gas Analyser

Serial number: IN12102

Analysers are calibrated on the morning of any pending CO2 calibrations in accordance with DJB Labcare procedures

Method

All calibrations are carried out in a temperature stable laboratory using reference gases providing traceability to UKAS ISO17025 standards and in accordance with the manufacturers procedures

Certified traceable sample gas from EffecTech, Laboratory Number 0590 Co2 5%
- Cylinder number D690633. Certificate Number: 18/0796/01

Date of calibration 3RD September 2019.

Approved by: Colin Bradnam

